


CERTIFICATE OF ANALYSIS No.: 2025-16069

CLIENT

KANNABIO HEMP HELLAS COOPERATIVE, SKOUFA 110 EL-38334 VOLOS, Greece

SAMPLE * LIVERAID

KANNABU



Sample condition:SUITABLEVSample ID:2505062ASample type:Viscous liquidMBatch No.:*KN0137M

 Work order:
 2025-112590

 Analysis ID:
 2025_031

 Method ID:
 PHL_RPC_16C

 Method SOP:
 MET-LAB-001-08

Sample received:31/01/2025Start of analysis:03/02/2025End of analysis:06/02/2025Analyst:Valentina Malin

* Information provided by the client.

CANNABINOID PROFILE		Concentration [% w/w]	Expanded uncertainty [% w/w]	Graphic presentation of relative cannabinoid concentration
CBDV	- Cannabidivarin	0.046	0.011	
CBDA	- Cannabidiolic acid	< LOQ	n/a	
CBGA	- Cannabigerolic acid	< LOQ	n/a	
CBG	- Cannabigerol	< LOQ	n/a	
CBD	- Cannabidiol	9.79	0.49	
THCV	- Tetrahydrocannabivarin	< LOQ	n/a	
CBN	- Cannabinol	< LOQ	n/a	
∆ ⁹ -THC	- Δ-9-Tetrahydrocannabinol	< LOQ	n/a	
∆ ⁸ -THC	- Δ-8-Tetrahydrocannabinol	< LOQ	n/a	
CBL	- Cannabicyclol	< LOQ	n/a	
СВС	- Cannabichromene	2.30	0.11	
∆ ⁹ -THCA	- Δ-9-Tetrahydrocannabinolic acid	< LOQ	n/a	
CBV	- Cannabivarin	< LOQ	n/a	
СВСА	- Cannabichromenic acid	< LOQ	n/a	
СВТ	- Cannabicitran	0.163	0.028	I
CBE	- Cannabielsoin	0.062	0.017	

Units and abbreviations: % w/w = weight percent, < LOQ = below the limit of quantitation (0.03 % w/w), ND = not detected, n/a = not available.

The results given herein apply only to the sample as received and tested. **Expanded Uncertainty** was calculated using coverage factor k = 2, corresponding to a double standard uncertainty and characterizes the interval value in which it is possible to expect the real value with a probability of 95%. This is stated according to the ISO/IEC Guide 98-3.

Total or partial reproduction of this document is not allowed without the permit from PharmaHemp d.o.o. The document does not substitute any other legal document.

Date issued:

Approved by:

06/02/2025

End of Certificate

mag. Valentina Malin Analytical Laboratory Manager Authorized by:

dr. Boštjan Jančar Chief Technology Officer